

WO#: 2201091

Date Reported: 1/14/2022

CLIENT: MAAS Corp

OMMA#: GAAA-JUMX-RXVT



Phone:

BatchID: 001

Sample Rec Date: 1/6/2022

Lot#:

LAB ID: 2201091-001A

Client Sample ID: Peach Ringz

Test	Result	Units	PQL	MDL	Qual	Date	Init	Max	OMMA
------	--------	-------	-----	-----	------	------	------	-----	------

POTENCY IN CANNABIS - BY HPLC

Cannabidiol (CBD)	< 0.225	wt%	0.225	0.225		1/11/2022	JR		
Cannabidiolic Acid (CBDA)	< 0.225	wt%	0.225	0.225		1/11/2022	JR		
Cannabigerol (CBG)	< 0.225	wt%	0.225	0.225		1/11/2022	JR		
Cannabigerolic Acid (CBGA)	1.54	wt%	0.225	0.225		1/11/2022	JR		
Cannabinol (CBN)	< 0.225	wt%	0.225	0.225		1/11/2022	JR		
Tetrahydrocannabinol (THC)	0.347	wt%	0.225	0.225		1/11/2022	JR		
Tetrahydrocannabinolic Acid (THCA)	19.0	wt%	0.225	0.225		1/11/2022	JR		
THC - Delta 8	< 0.225	wt%	0.225	0.225		1/11/2022	JR		
Total CBD (CBDAx0.877)+CBD	< 0.225	wt%	0.225	0.225		1/11/2022	JR		
Total CBG (CBGAx0.878)+CBG	1.35	wt%	0.225	0.225		1/11/2022	JR		
Total THC (THCAx0.877)+THC	17.0	wt%	0.225	0.225		1/11/2022	JR		

TERPENES IN CANNABIS - BY HS GC/FID

(-)-alpha-Bisabolol	227	ppm	43.5	43.5		1/10/2022	KP		
(-)-beta-Pinene	846	ppm	43.5	43.5		1/10/2022	KP		
(-)-Guaiol	436	ppm	43.5	43.5		1/10/2022	KP		
(-)-Isopulegol	< 43.5	ppm	43.5	43.5		1/10/2022	KP		
alpha-Humulene	3,170	ppm	43.5	43.5		1/10/2022	KP		
alpha-Ocimene	65.9	ppm	43.5	43.5		1/10/2022	KP		
alpha-Pinene	509	ppm	43.5	43.5		1/10/2022	KP		
alpha-Terpinene	< 43.5	ppm	43.5	43.5		1/10/2022	KP		
beta-Caryophyllene	8,670	ppm	43.5	43.5		1/10/2022	KP		
beta-Myrcene	2,680	ppm	43.5	43.5		1/10/2022	KP		
beta-Ocimene	< 43.5	ppm	43.5	43.5		1/10/2022	KP		
Camphene	205	ppm	43.5	43.5		1/10/2022	KP		
Caryophyllene oxide	102	ppm	43.5	43.5		1/10/2022	KP		
cis-Nerolidol	< 43.5	ppm	43.5	43.5		1/10/2022	KP		
delta-3-Carene	< 43.5	ppm	43.5	43.5		1/10/2022	KP		
D-Limonene	5,650	ppm	43.5	43.5		1/10/2022	KP		
1,8-Cineole	54.8	ppm	43.5	43.5		1/10/2022	KP		
gamma-Terpinene	< 43.5	ppm	43.5	43.5		1/10/2022	KP		
Geraniol	< 43.5	ppm	43.5	43.5		1/10/2022	KP		
Linalool	2,360	ppm	43.5	43.5		1/10/2022	KP		
p-Isopropyltoluene	< 43.5	ppm	43.5	43.5		1/10/2022	KP		
Terpinolene	74.1	ppm	43.5	43.5		1/10/2022	KP		
Total Terpene Percentage	2.52	%	0	0		1/10/2022	KP		
Total Terpenes	25,300	ppm	0	0		1/10/2022	KP		
trans-Nerolidol	201	ppm	43.5	43.5		1/10/2022	KP		

VISUAL INSPECTION OF CANNABIS - BY MICROSCOPE

1/4 area Foreign Material	< 1.0	Each	1.0	1.0		1/14/2022	BH	1.0	Pass
1/4 area Mold	< 1.0	Each	1.0	1.0		1/14/2022	BH	1.0	Pass
1/4 area Soil / Dirt / Sand	< 1.0	Each	1.0	1.0		1/14/2022	BH	1.0	Pass
Excreta, Mammalian	< 1.0	Each	1.0	1.0		1/14/2022	BH	1.0	Pass
Hair, Human	< 1.0	Each	1.0	1.0		1/14/2022	BH	1.0	Pass

LAB ID:

2201091-001A

Client Sample ID:

Peach Ringz

Test	Result	Units	PQL	MDL	Qual	Date	Init	Max	OMMA
<u>VISUAL INSPECTION OF CANNABIS - BY MICROSCOPE</u>									
Hair, Rodent	< 1.0	Each	1.0	1.0		1/14/2022	BH	1.0	Pass
Insect Fragment	< 1.0	Each	1.0	1.0		1/14/2022	BH	1.0	Pass
Filth & Foreign Matter	< 1.0	Each	1.0	1.0		1/14/2022	BH	1.0	Pass
<u>HEAVY METALS BY ICP/MS</u>									
Arsenic	< 0.155	ppm	0.155	0.155		1/12/2022	KR	0.200	Pass
Cadmium	< 0.155	ppm	0.155	0.155		1/12/2022	KR	0.200	Pass
Lead	< 0.388	ppm	0.388	0.388		1/12/2022	KR	0.500	Pass
Mercury	< 0.0776	ppm	0.0776	0.0776		1/12/2022	KR	0.100	Pass
<u>CANNABIS MICROBIALS - BY SPECTRAL ARRAY</u>									
Aspergillus flavus	< 1.00	CFU/g	1.00	1.00		1/12/2022	AC	1.00	Pass
Aspergillus fumigatus	< 1.00	CFU/g	1.00	1.00		1/12/2022	AC	1.00	Pass
Aspergillus niger	< 1.00	CFU/g	1.00	1.00		1/12/2022	AC	1.00	Pass
Aspergillus terreus	< 1.00	CFU/g	1.00	1.00		1/12/2022	AC	1.00	Pass
Salmonella	< 1.00	CFU/g	1.00	1.00		1/12/2022	AC	1.00	Pass
STEC E. coli	< 1.00	CFU/g	1.00	1.00		1/12/2022	AC	1.00	Pass
Yeast and Mold	2,200	CFU/g	1.00	1.00		1/12/2022	AC	10,000	Pass
<u>PERCENT MOISURE IN CANNABIS - BY MOISTURE BALANCE</u>									
Percent Moisture	9.92	%	0.0500	0.0500		1/10/2022	BH	15.0	Pass
<u>PESTICIDE RESIDUE IN CANNABIS - BY LC/MS/MS</u>									
Abamectin	< 0.469	ppm	0.469	0.469		1/12/2022	KP	0.500	Pass
Azoxystrobin	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Bifenazate	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Etoxazole	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Imazalil	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Imidacloprid	< 0.375	ppm	0.375	0.375		1/12/2022	KP	0.400	Pass
Malathion	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Myclobutanil	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Permethrin-cis	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Permethrin-trans	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Spinosyn A	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Spinosyn D	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Spiromesifen	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Spirotetramat	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Tebuconazole	< 0.375	ppm	0.375	0.375		1/12/2022	KP	0.400	Pass
Total Permethrin	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
Total Spinosad	< 0.188	ppm	0.188	0.188		1/12/2022	KP	0.200	Pass
<u>WATER ACTIVITY IN CANNABIS - BY WATER ACTIVITY</u>									
Water Activity	0.40	Aw	0.010	0.010		1/7/2022	BH	0.65	Pass

Comments:

QUALIFIER: * = Failure of analyte as per OMMA specifications
 OMMA Section: OMMA Exceedance limit and Pass/Fail section

The data and information presented herein, while not guaranteed, are to the best of our knowledge accurate and true. No warranty or guarantee implied or expressed is made regarding these analytical results, since securing and properly preserving representative samples and since sample custody chains are beyond GCT control. . Alteration or reproduction other than in its entirety is not authorized by Green Country Testing, Inc. Accreditation scope documents can be inspected at www.greencountrytesting.com or are available by request. A2LA certificate numbers 570.01 and 570.02. The recording of False, Fictitious or Fraudulent Statements or entry on this document may be punishable as a Felony under Federal Statute. All testing performed under GCT quality manual Rev 003. issued 6/21/21 and has been audited and deemed compliant to ISO Guide 17025 rev. 2005. GCT OMMA License Number: LAAA-AYUG-T2AU OBNDD: 61696


Approved By:

Brian Duzan
 Laboratory Director